

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Colcemid Lösung</b> sterilfiltriert	<b>Kat.Nr./Cat.No. :</b>	<b>47253</b>
Colcemid solution sterile filtered	<b>Lot/Contr.No. :</b>	<b>P240347</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare, farblose Lösung clear, colorless solution	entspricht corresponds
<b>Konzentration (µg/ml)</b> concentration	in PBS	10	entspricht corresponds
<b>pH</b>		7.0 – 8.0	7.3
<b>Osmolarität (mOsmol/l)</b> osmolarity		max. 300	281
<b>A 1cm/ 246 nm</b>	unverdünnt undiluted	min. 0.8	0.817
<b>A 1cm/ 354 nm</b>	unverdünnt undiluted	min. 0.4	0.408
<b>Sterilität</b> sterility		negativ negative	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			07.11.2026
<b>Lagerung (°C)</b> storage			-15 bis/ to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 13.11.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.