

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

Ethanol unvergällt absolut p.A.	Kat.Nr./Cat.No. : 11093
Ethanol undenaturated absolute analytical grade	Lot/Contr.No. : 230596

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₂ H ₆ O	
Molekulargewicht molecular weight		46.07	
Aussehen appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
Gehalt V/V (%) assay		min. 99.7	100
Wasser (%) water	KF	max. 0.2	0.01
Brechungsindex refractive index	n _D ²⁰	1.3600 – 1.3620	1.361
Mindesthaltbarkeit minimum shelf life			11.05.2028
Lagerung (°C) storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
 Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 23.05.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.