

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>Ampicillin-Na-salz, reinst, Ph. Eur.</b>	<b>Kat.Nr./Cat.No. :</b>	<b>13398</b>
Ampicillin-Na-salt, research grade, Ph. Eur.	<b>Lot/Contr.No. :</b>	<b>191201</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>16</sub> H <sub>18</sub> N <sub>3</sub> O <sub>4</sub> S·Na	
<b>Molekulargewicht</b> molecular weight		371.4	
<b>Aussehen</b> appearance		weißes bis fast weißes, hygroskopisches Pulver white or almost white powder, hygroscopic	entspricht corresponds
<b>Gehalt (%)</b> assay	getrocknet anhydrous	91.0 – 102.0	94.7
<b>Identität</b> identity	Ph. Eur.	entspricht corresponds	entspricht corresponds
<b>pH</b>	10% in H <sub>2</sub> O	8.0 – 10.0	9.1
<b>[α]<sub>D</sub>/ 20 °C</b>	Ph. Eur.	+258° bis/ to +287°	+277°
<b>Wasser (%)</b> water	KF	max. 2.0	0.8
<b>Mindesthaltbarkeit</b> minimum shelf life			25.11.2022
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die physikalisch-chemischen Eigenschaften entsprechen den Anforderungen der Ph. Eur.**  
 The physical/ chemical data corresponds to the requirements of Ph. Eur.

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 04.12.2019

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.