

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Renlam (R) M-1	Kat.Nr./Cat.No. : 13825
Renlam (R) M-1	Lot/Contr.No. : 140709

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description		klare, leicht gelbliche, viskose Flüssigkeit clear, slightly yellowish, viscous liquid	entspricht corresponds
Viskosität (mPa·s) viscosity	25 °C	1400 – 1800	1760
Epoxidzahl (Eq/kg) epoxy number		4.10 – 4.35	4.26
Dichte (g/cm³) density		1.10 – 1.15	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			04/ 2020
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 26.06.2014

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.