

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

| | |
|--|--------------------------------|
| SERVAColor BCIP/NBT Blot-Lösung | Kat.Nr./Cat.No. : 15245 |
| SERVAColor BCIP/NBT Blot Solution | Lot/Contr.No. : 220945 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|--|-------------------|------------------------------------|---------------------------|
| Aussehen appearance | | gelbe Flüssigkeit yellow liquid | entspricht corresponds |
| pH | | 9.5 – 9.9 | entspricht corresponds |
| visuelle Nachweisgrenze auf NC (ng/ml) detection limit (visual) at nitrocellulose membranes | | < 10 | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | | 14.08.2025 |
| Lagerung (°C) storage | | | +2 bis/ to +8 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 20.10.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.