

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>BCIP/NBT, gebrauchsfertiges Substrat für Immunhistochemie und Blotting</b>	<b>Kat.Nr./Cat.No. : 15246</b>
BCIP/NBT, ready-to-use substrate for Immunohistochemistry and Blotting	<b>Lot/Contr.No. : 240466</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		farblose bis schwach gelbliche Lösung colourless to pale yellow solution	entspricht corresponds
<b>pH</b>		9.7 – 9.9	entspricht corresponds
<b>Anwendung</b> application	Dot Blot	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			28.02.2027
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 27.05.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.