

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Cholesterin</b> reinst	<b>Kat.Nr./Cat.No. :</b>	<b>17101</b>
Cholesterol research grade	<b>Lot/Contr.No. :</b>	<b>130618</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>27</sub> H <sub>46</sub> O	
<b>Molekulargewicht</b> molecular weight		386.7	
<b>Aussehen</b> appearance		weißes Pulver white powder	entspricht corresponds
<b>Gehalt (%)</b> assay	GC	min. 95.0	97.3
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Schmelzpunkt (°C)</b> melting point		147 – 150	149
<b>Trocknungsverlust</b> loss on drying		max. 0.3	0.2
<b>Mindesthaltbarkeit</b> minimum shelf life			05/ 2018
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 28.05.2013**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.