

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

Glutaraldehyd 25% in Wasser Standardqualität	Kat.Nr./Cat.No. :	23115
Glutaraldehyde 25% in water standard grade	Lot/Contr.No. :	220651

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₅ H ₈ O ₂	
Molekulargewicht molecular weight		100.13	
Aussehen appearance		klare, farblose bis gelbliche Lösung clear, colorless to yellowish solution	entspricht corresponds
Gehalt (%) assay		23.0 – 27.0	25.5
Brechungsindex refractive index	25 °C	1.3690 – 1.3755	1.3734
pH	25 % in H ₂ O	3.0 – 4.0	3.4
Mindesthaltbarkeit minimum shelf life			30.06.2024
Lagerung (°C) storage			+2 bis/ to +8

<p>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert. Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle. We do not guarantee that the product can be used for a special application. This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 11.07.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.