

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Glyceringelatine für Mikroskopie</b>	<b>Kat.Nr./Cat.No. :</b>	<b>23310</b>
Glycerol gelatin for microscopy	<b>Lot/Contr.No. :</b>	<b>130396</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance	gelbbraune Masse yellowish brown solid	entspricht corresponds
<b>Löslichkeit in Wasser (50°C)</b> solubility in water	löslich soluble	entspricht corresponds
<b>Eignung für Mikroskopie</b> suitability for microscopy		entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		03/ 2015
<b>Lagerung (°C)</b> storage		+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 10.04.2013**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.