

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

N,N'-Methylenbisacrylamid 2 x Lösung 2 % (w/v)	Kat.Nr./Cat.No. : 29197
N,N'-Methylene bisacrylamide 2 x Lösung 2 % (w/v)	Lot/Contr.No. : P220098

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Lösung enthält 2 % (w/v) hochreines „Bis“ in deionisiertem Wasser Solution contains 2 % (w/v) highly purified „Bis“ in deionized water		
Aussehen appearance		klare farblose Lösung clear colorless solution	entspricht corresponds
A 1 cm/ 290 nm	1:1 in H ₂ O	max. 0.25	0.195
pH	1:1 in H ₂ O	6.0 – 8.0	6.2
Polymerisationszeit (min) polymerizing time	15 % Gel, 25 °C	max. 30	entspricht corresponds
Mindesthaltbarkeit minimum shelf life	16.09.2023		
Lagerung (°C) storage	+2 bis/ to +8		

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 28.03.2022

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.