

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Osmium tetroxid</b> für Elektronenmikroskopie	<b>Kat.Nr./Cat.No. :</b>	<b>31251</b>
Osmium tetroxide for electron microscopy	<b>Lot/Contr.No. :</b>	<b>130980</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		OsO <sub>4</sub>	
<b>Molekulargewicht</b> molecular weight		254.2	
<b>Aussehen</b> appearance		gelbe bis gelbgrüne Kristalle yellow to greenish crystals	entspricht corresponds
<b>Gehalt (%)</b> assay	OsO <sub>4</sub> Os	min. 99.9	99.9 75
<b>nicht flüchtige Anteile (%)</b> non volatile content		< 0.02	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			08/ 2018
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 28.08.2013**

Dipl.-Ing. (FH) Bernhard Göckel

Jakob Arndt

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.