

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Propidiumjodid</b> reinst	<b>Kat.Nr./Cat.No. :</b>	<b>33671</b>
Propidium iodide research grade	<b>Lot/Contr.No. :</b>	<b>160827</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Summenformel</b> formula		$C_{27}H_{34}N_4^{2+} \cdot 2I^-$	
<b>Molekulargewicht</b> molecular weight		668.4	
<b>Aussehen</b> appearance		feines, weinrotes Pulver fine, dark-red powder	entspricht corresponds
<b>Reinheit (%)</b> purity	HPLC	min. 90.0	98.4
<b>Mindesthaltbarkeit</b> minimum shelf life			30.11.2018
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 12.09.2017

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.