

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Propidiumjodid</b> reinst	<b>Kat.Nr./Cat.No. :</b>	<b>33671</b>
Propidium iodide research grade	<b>Lot/Contr.No. :</b>	<b>180259</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		$C_{27}H_{34}N_4^{2+} \cdot 2I^-$	
<b>Molekulargewicht</b> molecular weight		668.40	
<b>Aussehen</b> appearance		rotes Pulver red powder	entspricht corresponds
<b>Gehalt (%)</b> assay	Titration	min. 90.0	95.42
<b>Mindesthaltbarkeit</b> minimum shelf life			01.07.2019
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 15.03.2018

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.