

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Silicon-Antischaum-Emulsion, 30%</b>	<b>Kat.Nr./Cat.No. : 35119</b>
Silicone antifoam emulsion, 30%	<b>Lot/Contr.No. : 150078</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		weiße Emulsion white emulsion	entspricht corresponds
<b>Gehalt (%)</b> assay	Polydimethylsiloxane	28.0 – 32.0	28.7
<b>pH</b>		2.0 – 3.0	2.8
<b>Antischaumaktivität (s)</b> defoaming activity	USP	max. 15	1
<b>Schwermetalle (ppm)</b> heavy metals	USP	max. 5	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			10/ 2017
<b>Lagerung (°C)</b> storage			+15 bis / to +30

**Dieses Produkt enthält keine Bestandteile tierischen Ursprungs.**

This product does not contain any components of animal origin.

**Die physikalisch-chemischen Eigenschaften entsprechen den Anforderungen der USP.**

The physical/ chemical data correspond to the requirements of USP.

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**

**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 20.01.2015**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**

This report has been computer generated and does not contain a signature.