

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Siliconöl DC 200; 10 cst pract.</b>	<b>Kat.Nr./Cat.No. : 35132</b>
Silicone DC 200 fluid; 10 cst, pract.	<b>Lot/Contr.No. : 140298</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Dichte (g/ml)</b> specific gravity	1013 hPa/ 25 °C	0.934 - 0.940	entspricht corresponds
<b>Brechungsindex</b> refractive index	25 °C	1.3980 - 1.4000	1.3990
<b>Viskosität (cSt)</b> viscosity		9.0 - 11.0	10.0
<b>Flüchtige Bestandteile (%)</b> volatile content		max. 1.0	0.12
<b>Mindesthaltbarkeit</b> minimum shelf life			03/ 2016
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 10.03.2014**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.