

ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS



<b>5-Sulfosalicylsäure p.A.</b>	<b>Kat.Nr./Cat.No. : 35706</b>
<b>5-Sulfosalicylic acid analytical grade</b>	<b>Lot/Contr.No. : 240522</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>7</sub> H <sub>6</sub> O <sub>6</sub> S·2H <sub>2</sub> O	
<b>Molekulargewicht</b> molecular weight		254.23	
<b>Aussehen</b> appearance		weiße Kristalle white crystals	entspricht corresponds
<b>Gehalt (%)</b> assay		min. 99,0	99.2
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Wasser (%)</b> water	KF	max. 22	14 - 16
<b>Mindesthaltbarkeit</b> minimum shelf life			10.06.2029
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 11.06.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.