

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

3,3',5,5'-Tetramethylbenzidin reinst	Kat.Nr./Cat.No. :	35926
3,3',5,5'-Tetramethylbenzidine research grade	Lot/Contr.No. :	130006

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₁₆ H ₂₀ N ₂	
Molekulargewicht molecular weight		240.4	
Aussehen appearance		hell elfenbeinfarbenes Pulver light ivory-colored powder	entspricht corresponds
Gehalt (%) assay	Titration	98.0 – 102.0	99.8
Schmelzpunkt (°C) melting point		168 – 170	169 – 170
Trocknungsverlust (%) loss on drying		max. 0.5	0.1
Identität identity	IR	entspricht corresponds	entspricht corresponds
Eisen (ppm) lead	ICP-OES		3
Mindesthaltbarkeit minimum shelf life			12/2015
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 24.09.2014

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.