

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Trichloressigsäure, 20 % Lösung	Kat.Nr./Cat.No. : 36913
Trichloroacetic acid, 20 % solution	Lot/Contr.No. : P150207

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
Gehalt TCA (g/l) assay TCA	200.0	entspricht corresponds
pH	< 0.5	0.42
Mindesthaltbarkeit minimum shelf life		06/ 2017
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 11.06.2015

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.