

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

---

<b>Trichloressigsäure, 20 % Lösung</b>	<b>Kat.Nr./Cat.No. :</b>	<b>36913</b>
Trichloroacetic acid, 20 % solution	<b>Lot/Contr.No. :</b>	<b>P250045</b>

---

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
<b>Gehalt TCA (g/l)</b> assay TCA	200.0	entspricht corresponds
<b>pH</b>	< 0.5	0.34
<b>Mindesthaltbarkeit</b> minimum shelf life		30.01.2027
<b>Lagerung (°C)</b> storage		+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

---

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control****Druckdatum/ printing date:** 06.02.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.