

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | |
|--|--------------------------------|
| TMB, gebrauchsfertiges ELISA-Substrat | Kat.Nr./Cat.No. : 37068 |
| TMB, ready-to-use ELISA substrate | Lot/Contr.No. : 170870 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|--------------------------|--|---------------------------|
| Aussehen appearance | | klare farblose bis leicht blaue Lösung clear colourless to slightly blue solution | entspricht corresponds |
| pH | | 3.75 – 3.85 | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | | 30.09.2018 |
| Lagerung (°C) storage | | | +2 bis/ to +8 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 13.09.2017

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.