

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

TMB, gebrauchsfertiges Substrat zum Blotten	Kat.Nr./Cat.No. : 37070
TMB, ready-to-use substrate for Blotting	Lot/Contr.No. : 160158

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose Lösung clear, colourless solution	entspricht corresponds
pH		4.87 – 4.93	entspricht corresponds
Anwendung application	Immunoblot	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			30.04.2017
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 17.02.2016

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.