

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>SERVAColor TMB Blot-Lösung</b>	<b>Kat.Nr./Cat.No. : 37071</b>
SERVAColor TMB Blot Solution	<b>Lot/Contr.No. : 170175</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		farblose bis leicht gelbe Flüssigkeit colourless to pale yellow liquid	entspricht corresponds
<b>pH</b>		4.9 – 5.1	5.0
<b>Mindesthaltbarkeit</b> minimum shelf life			30.09.2020
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

SERVA Electrophoresis GmbH

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 22.10.2019**

Christian Monsler

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.