

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

---

<b>Tergitol™ 15-S-9</b>	<b>Kat.Nr./Cat.No. :</b> 37242
	<b>Lot/Contr.No. :</b> 220370

---

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		farblose bis gelbe, viskose Flüssigkeit colourless to yellow, viscous liquid	entspricht corresponds
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Wasser (%)</b> water	KF	max. 0.5	0.06
<b>pH</b>	1% in H <sub>2</sub> O	5.5 – 7.5	6.0
<b>Trübungspunkt (°C)</b> cloud point	1% in H <sub>2</sub> O	57.5 – 62.5	58.3
<b>Ursprung</b> origin		nicht tierisch non animal	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			01.12.2025
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

---

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 14.04.2022

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.