

CERTIFICATE OF ANALYSIS



Protein Test Mixture 4

Cat.No. : 39208

Lot No.: P240377

| Protein Components | Molecular weight |
|---------------------------|-------------------------|
|---------------------------|-------------------------|

| | |
|---------------------------|--------|
| Carbonic anhydrase | 29 000 |
|---------------------------|--------|

| | |
|--------------------|--------|
| Albumin egg | 45 000 |
|--------------------|--------|

| | |
|-----------------------|--------|
| Albumin bovine | 67 000 |
|-----------------------|--------|

| | |
|------------------------|--------|
| Phosphorylase B | 97 400 |
|------------------------|--------|

| | |
|---------------------------|------------|
| Minimum shelf life | 28.11.2026 |
|---------------------------|------------|

| | |
|---------------------|----------|
| Storage (°C) | +2 to +8 |
|---------------------|----------|

Reconstitution Dissolve to a concentration of 1mg/ml in SDS-sample buffer. A volume of 5µl is sufficient for SDS-PAGE (in a standard form of 80 x 60 x 0.75mm).

SDS-sample buffer:

125 mM Tris-HCl pH 6.8; 2% SDS; 15% glycerol;
10 mM DTT; 0.025% Bromphenol blue; 0.025% Orange G

**We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.**

SERVA Electrophoresis GmbH
Quality Control

Printing date: 29.11.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

This report has been computer generated and does not contain a signature.