

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Bradford-Reagenz, 5x Konzentrat	Kat.Nr./Cat.No. :	39222
Bradford reagent, 5x	Lot/Contr.No. :	P150085

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		rote bis braune Lösung red to brown solution	entspricht corresponds
Anwendung application	Bradford assay	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			03/ 2016
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 03.03.2015

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.