

**ANALYSENZERTIFIKAT**  
 CERTIFICATE OF ANALYSIS

<b>Isopropanol</b>	<b>Kat.Nr./Cat.No. :</b>	<b>39559</b>
molecular biology grade	<b>Lot/ Contr.No. :</b>	<b>250141</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>3</sub> H <sub>8</sub> O	
<b>Molekulargewicht</b> molecular weight		60.09	
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colourless liquid	entspricht corresponds
<b>Gehalt (%)</b> assay	GC	min. 99.8	100.0
<b>Dichte (g/ml)</b> specific density	20 °C	0.784 – 0.786	0.786
<b>Brechungsindex</b> refractive index	20 °C	1.376 – 1.379	1.3771
<b>Wasser (%)</b> water		max. 0.15	0.14
<b>DNase/ RNase Aktivität</b> DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			15.01.2030
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**
**Druckdatum/ printing date: 17.02.2025**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.