

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Ammoniumacetat-Lösung 7.5M</b> Ammonium acetate solution 7.5M molecular biology grade	<b>Kat.Nr./Cat.No. : 39751</b> <b>Lot/Contr.No. : P170001</b>
--	--

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare farblose Lösung clear colorless solution	entspricht corresponds
<b>Zusammensetzung</b> composition			
Ammoniumacetat (#39750) ammonium acetate		578.1 g/l	entspricht corresponds
<b>Schwermetalle (%)</b> heavy metals	als/ as Pb	max. 0.001	entspricht corresponds
<b>DNase/ RNase Aktivität</b> DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			02.01.2019
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 17.01.2017**

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.