

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | | |
|---|--------------------------|----------------|
| Magnesiumchloridlösung, 1M | Kat.Nr./Cat.No. : | 39772 |
| Magnesium chloride, solution 1M molecular biology grade | Lot/Contr.No. : | P160182 |

| Parameter parameter | Spezifikation specification | Ergebnis result |
|--|---|---------------------------|
| Aussehen appearance | klare, farblose Lösung clear, colorless solution | entspricht corresponds |
| pH | 4.5 – 7.0 | 5.8 |
| Zusammensetzung composition | | |
| MgCl ₂ · 6H ₂ O (#39771) | 203.30 g/l | entspricht corresponds |
| DNase/ RNase Aktivität DNase/ RNase activity | nicht nachweisbar none detected | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | 28.04.2019 |
| Lagerung (°C) storage | | +15 bis/ to +30 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control**Druckdatum/ printing date: 20.02.2018**

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.