

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Tris-Puffer pH 8.0, 1M Lösung</b>	<b>Kat.Nr./Cat.No. : 39792</b>
Tris buffer pH 8.0, solution 1M molecular biology grade	<b>Lot/Contr.No. : P240139</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare farblose Lösung clear colorless solution	entspricht corresponds
<b>Zusammensetzung</b> composition			
TRIS (#37186)		121.14 g/l	entspricht corresponds
<b>pH</b>	20°C eingestellt mit HCl adjusted with HCl	7.9 - 8.1	8.1
<b>DNase/ RNase Aktivität</b> DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			18.04.2026
<b>Lagerung (°C)</b> storage			+15 bis / to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 22.04.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.