

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

CTAB DNA-Extraktionspuffer , Molekularbiologie-Qualität	Kat.Nr./Cat.No. : 39809
CTAB DNA Extraction Buffer, molecular biological grade	Lot/Contr.No. : P190212

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare bis opaleszierende, farblose Lösung clear to opalescent, colourless solution	entspricht corresponds
pH	7.8 – 8.2	7.9
DNase/ RNase Aktivität DNase/ RNase activity	nicht nachweisbar none detected	entspricht corresponds
Zusammensetzung composition		
CTAB (# 16530)	20 g/l	entspricht/ corresponds
TRIS•HCl (# 37192)	0.1 mol/l	entspricht/ corresponds
EDTA•Na ₂ •2H ₂ O (# 11280)	20 mmol/l	entspricht/ corresponds
NaCl (# 30183)	1.4 mol/l	entspricht/ corresponds
Mindesthaltbarkeit minimum shelf life		13.06.2022
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert. Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle. We do not guarantee that the product can be used for a special application. This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 17.06.2019

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.