

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DOWEX® 1X2 (50 – 100 mesh) pract.	Kat.Nr./Cat.No. : 41010
	Lot/Contr.No. : 100849

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		reinweißes bis elfenbeinfarbenes Korn off-white to ivory- colored particles	entspricht corresponds
Trocknungsverlust (%) loss on drying	3 h, 105°C	65.0 – 75.0	65.0
Gegenion counter-ion		Cl ⁻	entspricht corresponds
Kapazität (eq/l) capacity		min. 0.7	0.7
Mindesthaltbarkeit minimum shelf life			03/ 2013
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 20.12.2011

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.