

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Dowex® 1X8 (50 – 100 mesh) , pract.	Kat.Nr./Cat.No.: 41090 Lot/Contr.No.: 111664
---	---

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		elfenbeinfarbenes Korn ivory colored particles	entspricht corresponds
Gegenion counter ion		Cl ⁻	entspricht corresponds
Trocknungsverlust (%) loss on drying	105 °C, 3 h	43.0 – 47.0	46.3
Kapazität (eq/l) capacity		min. 1.2	1.6
Mindesthaltbarkeit minimum shelf life			25.10.2018
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 26.07.2016

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.