

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Dowex® 1X8</b> (50 – 100 mesh) , pract.	<b>Kat.Nr./Cat.No.:</b> 41090 <b>Lot/Contr.No.:</b> 121041
---	---

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		elfenbeinfarbenes Korn ivory colored particles	entspricht corresponds
<b>Gegenion</b> counter ion		Cl <sup>-</sup>	entspricht corresponds
<b>Trocknungsverlust (%)</b> loss on drying	105 °C, 3 h	43.0 – 47.0	45.4
<b>Kapazität (eq/l)</b> capacity		min. 1.2	1.4
<b>Mindesthaltbarkeit</b> minimum shelf life			09/ 2017
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 25.09.2012

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.