

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>DOWEX® 50 WX8</b>	<b>Kat.Nr./Cat.No. : 41620</b>
(100 – 200 mesh) pract.	<b>Lot/Contr.No. : 121329</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		beige-grünes Korn beige-green-colored particles	entspricht corresponds
<b>Trocknungsverlust (%)</b> loss on drying	3 h, 105 °C	50.0 – 58.0	51.4
<b>Gegenion</b> counter-ion		H <sup>+</sup>	entspricht corresponds
<b>Kapazität (eq/l)</b> capacity		min. 1.7	2.1
<b>Mindesthaltbarkeit</b> minimum shelf life			08/ 2015
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 15.05.2014**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.