

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

DOWEX® 50 WX8 p.A.	Kat.Nr./Cat.No. : 41621
(100 – 200 mesh) analytical grade	Lot/Contr.No. : P250186

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		beiges Korn beige-colored particles	entspricht corresponds
Trocknungsverlust (%) loss on drying	3 h, 105 °C	45.0 – 55.0	45.5
Gegenion counter-ion		Na ⁺	entspricht corresponds
Kapazität (eq/l) capacity		min. 1.7	2.0
Mindesthaltbarkeit minimum shelf life			23.05.2030
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 27.05.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.