

**A N A L Y S E N Z E R T I F I K A T**  
**C E R T I F I C A T E O F A N A L Y S I S**

<b>Serdolit<sup>®</sup> PAD-II p.A.</b> (0.3 – 1.0mm) analytical grade	<b>Kat.Nr./Cat.No.:</b> 42445 <b>Lot/Contr.No.:</b> 160884
---	---

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		hellelfenbeinfarbenes Korn light ivory-colored particles	entspricht corresponds
<b>Zusammensetzung</b> composition		Ethylstyrol-DVB-Kopolymer ethylstyrene-DVB-copolymer	entspricht corresponds
<b>Trocknungsverlust (%)</b> loss on drying	3h; 105°C	30.0 – 40.0	34.6
<b>Spez. Oberfläche (m<sup>2</sup>/g)</b> specific surface		min. 400	522
<b>Mindesthaltbarkeit</b> minimum shelf life			31.07.2021
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 20.09.2016

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.