

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

Serdolit® PAD-III p.A. (0.1 – 0.2mm) analytical grade	Kat.Nr./Cat.No.: 42449 Lot/Contr.No.: 120403
---	---

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Zusammensetzung composition		Ethylstyrol-DVB-Kopolymer ethylstyrene-DVB-copolymer	entspricht corresponds
Aussehen appearance		rotbraunes feines Korn redbrown fine particles	entspricht corresponds
Trocknungsverlust (%) loss on drying	3 h; 105 °C	45.0 – 55.0	49.2
Spez. Oberfläche (m²/ g) specific surface		min. 800	944
Porendurchmesser (nm) pore diameter		ca. 25	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			02/ 2017
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 29.03.2012

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.