

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVA Tris-Glycin nativer Elektrophorese Puffer (10x)	Kat.Nr./ Cat.No. : 42530
SERVA Tris-Glycine native electrophoresis buffer (10x)	Lot/ Contr.No. : P250240

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	8.4 – 8.8	8.7
Zusammensetzung composition		
TRIS (# 37180)	0.25 M	entspricht corresponds
Glycin/ Glycine (# 23390)	1.92 M	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		23.01.2027
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 28.07.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.