

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Nativer Anodenpuffer für Blue/ Clear Native (10x)	Kat.Nr./Cat.No. :	42535
Native Anode Buffer for Blue/ Clear Native (10x)	Lot/Contr.No. :	P240238

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
pH	6.9 – 7.1	7.0
Zusammensetzung composition		
2-[Bis(2-hydroxyethyl)amino]-2-(hydroxy- methyl)- 1,3-propanediol / BISTRIS (#15107)	500 mM	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		18.07.2026
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 06.08.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.