

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Nativer Anodenpuffer für Blue/ Clear Native (10x)</b>	<b>Kat.Nr./Cat.No. : 42535</b>
Native Anode Buffer for Blue/ Clear Native (10x)	<b>Lot/Contr.No. : P240389</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
<b>pH</b>	6.9 – 7.1	7.0
<b>Zusammensetzung</b> composition		
2-[Bis(2-hydroxyethyl)amino]-2-(hydroxy- methyl)- 1,3-propanediol / BISTRIS (#15107)	500 mM	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		13.12.2026
<b>Lagerung (°C)</b> storage		+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 17.12.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.