

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

Nativer Kathodenpuffer für Blue/ Clear Native (10x)	Kat.Nr./Cat.No. :	42536
Native Cathode Buffer for Blue/ Clear Native (10x)	Lot/Contr.No. :	P180085

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
pH	6.9 – 7.1	7.1
Zusammensetzung composition		
N-Tris(hydroxymethyl)methylglycin (#37195)	500 mM	entspricht corresponds
Bis(2-hydroxyethyl)amino]-2-(hydroxy- methyl)- 1,3-propanediol/ BISTRIS (#15107)	150 mM	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		12.03.2020
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 27.03.2018

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.