

ANALYSENZERTIFIKAT
 CERTIFICATE OF ANALYSIS

TAE Puffer (50x)	Kat.Nr./Cat.No. : 42549
TAE Buffer (50x) molecular biology grade	Lot/Contr.No. : P190409

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
pH	8.1 – 8.5	8.5
Zusammensetzung composition		
TRIS (# 37180)	242.3 g/l	entspricht corresponds
EDTA-Na ₂ ·2H ₂ O (# 11280)	18.6 g/l	entspricht corresponds
Eisessig/ Acetic acid	60.05 g/l	entspricht corresponds
DNase/ RNase Aktivität DNase/ RNase activity	nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		13.12.2021
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control
Druckdatum/ printing date: 16.12.2019

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.