

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|                         |                                |
|-------------------------|--------------------------------|
| <b>TAE Puffer (10x)</b> | <b>Kat.Nr./Cat.No. : 42553</b> |
| TAE buffer (10x)        | <b>Lot/Contr.No. : P120475</b> |

| <b>Parameter</b><br>parameter                     | <b>Spezifikation</b><br>specification               | <b>Ergebnis</b><br>result |
|---|---|---------------------------|
| <b>Aussehen</b><br>appearance                     | klare, farblose Lösung<br>clear, colorless solution | entspricht<br>corresponds |
| <b>pH</b>   | 8.3 – 8.7   | 8.3                       |
| <b>Zusammensetzung</b><br>composition             |   |                           |
| TRIS (# 37180)                                    | 0.4 M   | entspricht<br>corresponds |
| EDTA-Na <sub>2</sub> ·2H <sub>2</sub> O (# 11280) | 0.01 M  | entspricht<br>corresponds |
| Eisessig<br>glacial acetic acid                   | 0.2 M   | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life   |   | 05/ 2015                  |
| <b>Lagerung (°C)</b><br>storage                   |   | +15 bis/ to +30           |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 18.03.2014**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.