

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SSC-Puffer (20x)	Kat.Nr./Cat.No. : 42555
SSC-buffer (20x) molecular biology grade	Lot/Contr.No. : P170284

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	6.8 - 7.2	7.1
Zusammensetzung composition		
NaCl (# 30183)	3.0 M	entspricht corresponds
tri-Natriumcitrat·2H ₂ O tri-Sodiumcitrate·2H ₂ O (# 38642)	0.3 M	entspricht corresponds
DNase/ RNase Aktivität DNase/ RNase activity	nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		21.02.2019
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 24.08.2017

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.