

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

„Laemmli“ Puffer 10x	Kat.Nr./Cat.No. : 42556
„Laemmli“ Buffer 10x	Lot/Contr.No. : P130478

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	8.4 – 8.9	8.7

Zusammensetzung
composition

TRIS (# 37180)	0.25 M	entspricht corresponds
Glycin / Glycine (# 23390)	1.92 M	entspricht corresponds
SDS (# 20763)	1 %	entspricht corresponds

Mindesthaltbarkeit
minimum shelf life

12/ 2014

Lagerung (°C)
storage

+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 28.11.2013

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.