

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

| | |
|----------------------|--------------------------------|
| „Laemmli“ Puffer 10x | Kat.Nr./Cat.No. : 42556 |
| „Laemmli“ Buffer 10x | Lot/Contr.No. : P220027 |

| Parameter parameter | Spezifikation specification | Ergebnis result |
|---|---|---------------------------|
| Aussehen appearance | klare, farblose Lösung clear, colorless solution | entspricht corresponds |
| pH | 8.4 – 8.9 | 8.8 |
| Zusammensetzung composition | | |
| TRIS (# 37180) | 0.25 M | entspricht corresponds |
| Glycin / Glycine (# 23390) | 1.92 M | entspricht corresponds |
| SDS (# 20770) | 1 % | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | 31.01.2024 |
| Lagerung (°C) storage | | +15 bis/ to +30 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 03.02.2022

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.