

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

|                      |                                |
|----------------------|--------------------------------|
| „Laemmli“ Puffer 10x | <b>Kat.Nr./Cat.No. : 42556</b> |
| „Laemmli“ Buffer 10x | <b>Lot/Contr.No. : P250049</b> |

| <b>Parameter</b><br>parameter                   | <b>Spezifikation</b><br>specification               | <b>Ergebnis</b><br>result |
|-------------------------------------------------|-----------------------------------------------------|---------------------------|
| <b>Aussehen</b><br>appearance                   | klare, farblose Lösung<br>clear, colorless solution | entspricht<br>corresponds |
| <b>pH</b>                                       | 8.4 – 8.9                                           | 8.8                       |
| <b>Zusammensetzung</b><br>composition           |                                                     |                           |
| TRIS (# 37180)                                  | 0.25 M                                              | entspricht<br>corresponds |
| Glycin / Glycine (# 23390)                      | 1.92 M                                              | entspricht<br>corresponds |
| SDS (# 20765)                                   | 1 %                                                 | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |                                                     | 30.01.2027                |
| <b>Lagerung (°C)</b><br>storage                 |                                                     | +15 bis/ to +30           |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 10.02.2025**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.