

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

TBE Puffer 10x	Kat.Nr./Cat.No. : 42557
TBE Buffer 10x	Lot/Contr.No. : P150077

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	8.1 – 8.5	8.3
Zusammensetzung composition		
TRIS (# 37180)	0.89 M	entspricht corresponds
EDTA-Na ₂ ·2H ₂ O (# 11280)	0.02 M	entspricht corresponds
Borsäure / Boric acid (# 15165)	0.89 M	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		02/ 2017
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 27.02.2015

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.