

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

SERVALight Vega	Kat.Nr./Cat.No. :	42588
CL HRP WB Substrate Kit	Lot/Contr.No. :	220642-220643

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose Lösung clear, colourless solution	entspricht corresponds
Anwendung application	Chemilumineszenz chemiluminescence	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			20.06.2024
Lagerung (°C) storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
 Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 14.07.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.