

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

PBS Puffer (10x), steril	Kat.Nr./Cat.No. : 42595
PBS Buffer (10x), sterile	Lot/Contr.No. : P210311

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose Lösung clear, colourless solution	entspricht corresponds
pH	1:10 verdünnt 1:10 diluted	7.2 – 7.6	7.4
Steriltest sterile test		kein Wachstum no growth	entspricht corresponds
Zusammensetzung composition			
NaCl (# 30183)		1.37 M	
KCl (# 26868)		27 mM	
Na ₂ HPO ₄ •2H ₂ O (# 30200)		100 mM	
KH ₂ PO ₄ (# 26870)		20 mM	
Mindesthaltbarkeit minimum shelf life			16.09.2023
Lagerung (°C) storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
 Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 27.09.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.